



Scope of Accreditation for Testing

As per ISO/IEC 17025

CAB Name : Greenglobe Fuel Solutions Pvt. Ltd. Testing and Calibration
Laboratory Division
Address : Greenhouse I, II & III, Building No. 5 & 6, Radhekrishna Industrial
Park, Pimplas Village, Mumbai Nashik Road, Bhiwandi, Thane –
421302, Maharashtra.
Certificate No.: C-0021

Page No. 1 of 1

Issue date: 06.08.2026

Validity: 05.08.2030

Sr. No.	Parameter/Measurand quantity, Instrument or gauge	Range	Remarks/Method used	CMC	Facility
Discipline: Fluid Flow Group: Flow Measuring Device (Gas Flow Meter)					
1	Gas Flow Meter	2.5 m ³ /hr (Q _{max})	Comparison method based on EN 1359:2017 and OIML R137:2014 using Sonic Nozzles as standard and flow test bench	1.5%	Permanent
2	Gas Flow Meter	1.75 m ³ /hr (0.7Q _{max})		1.5%	Permanent
3	Gas Flow Meter	1.0 m ³ /hr (0.4Q _{max})		1.5%	Permanent
4	Gas Flow Meter	0.5 m ³ /hr (0.2Q _{max})		1.5%	Permanent
5	Gas Flow Meter	0.25 m ³ /hr (0.1Q _{max})		1.5%	Permanent
6	Gas Flow Meter	0.046 m ³ /hr (3Q _{min})		3.0%	Permanent
7	Gas Flow Meter	0.016 m ³ /hr (Q _{min})		3.8%	Permanent

Note: CMC in (±) at 95% confidence level



Certificate

